



DIACHEM Ltd

H-1117 Budapest, Budafoki út 111-113

Phone: +36-1-205-3246,

+36-1-205-3247

Phone/Fax: +36-1-205-3616

CHLORIDE

LIQUID



Cat. No.: 41641
2x125 ml

41643
20x20 ml



Reagent kit for determination of chloride ion concentration in serum, urine and cerebrospinal fluid.
A colorimetric endpoint method based on the reaction with mercuric thiocyanate.

Chloride ion represents that anion of the salt and water household which is present in the highest concentrations in the body. It is primarily found as NaCl in the extracellular compartment and as HCl in the gastrointestinal tract. Low chloride concentrations are associated with severe vomiting, diarrhoea, colitis ulcerosa, diabetic acidosis, Addison's disease. Decreased chloride concentrations are found also in cases when drugs that need chloride ions for their absorption are taken for prolonged time.

Increased levels are observed in cases of dehydration, congestive heart failure, Cushing's syndrome, hyperventilation, anaemia, nephritis and renal obstruction.

Principle

Chloride ion in acidic environment in presence of ferric nitrate forms a colored complex with mercuric thiocyanate. Intensity of the developed colour is proportional to the chloride ion concentration in the sample.



Reference values

Serum: 98-107 mmol/l (348-380 mg/dl)

Cerebrospinal fluid: 119-131 mmol/l (421-465 mg/dl)

Urine: 110-250 mmol/24 h (389-887 mg/24 h)

It is recommended that each laboratory should assign its own normal range.

Reagents

1. Reagent (R1)

Mercuric thiocyanate 2 mmol/l
Ferric nitrate 20 mmol/l
Nitric acid 29 mmol/l

2. Reagent (R2)

Chloride standard See label for exact value.
20x20 ml kit doesn't contain any standard.

Samples

Serum, urine, cerebrospinal fluid.

PROCEDURE

Reagent is ready for use.

If the absorbance of working reagent is higher than 0.1 at 492 nm the reagent can not be used.

Assay conditions

Wavelength: 500 (480-520) nm
Temperature: 37°C
Cuvette: 1 cm light path
Method: endpoint (increasing)

Pipette into cuvette

	Blank	Standard	Sample
Reagent	1 ml	1 ml	1 ml
Distilled water	10 µl		
Standard		10 µl	
Sample			10 µl

Mix and read the absorbance against the blank.

Calibration: (37°C, mercuric thiocyanate method)

S1: Distilled water

S2: Chloride standard found in the kit or

Roche C.F.A.S. (Calibrator for automated system) or

Randox Calibration Serum Level I or

Randox Calibration Serum Level II

Calibration frequency

Two point calibration is recommended:

- after reagent lot change,

- as required following quality control procedures.

Calculation using calibration

$$\frac{A_{\text{sample}}}{A_{\text{standard}}} \times C_{\text{standard}} = C_{\text{sample}}$$

A = absorbance, C = concentration

Quality control

Quality control program is recommended for all clinical laboratories. The analysis of control material in both the normal and abnormal ranges with each assay is recommended for monitoring the performance of the procedure. Each laboratory should establish corrective measures to be taken if values fall outside the limits.

PERFORMANCES DATA

The following data were obtained using the Hitachi 717 analyzer (37°C).

Linearity

Relationship of absorbance vs concentration is linear up to chloride ion concentration of 135 mmol/l (479 mg/dl)

Sensitivity

It is recommended that each laboratory establishes its own range of sensitivity as this is limited by the sensitivity of the spectrophotometer used. Under manual conditions however, a change of 0.001 Abs is equivalent to 0.25 mmol/l (0,89 mg/dl) chloride concentration at 492 nm.

Precision

Reproducibility			
Sample	Average concentration (mmol/l)	SD	CV%
sample I	111	2.6	2.35
sample II	109	1.3	1.19

Repeatability			
Sample	Average concentration (mmol/l)	SD	CV%
sample I	83.4	1.19	1.42
sample II	112	1.05	0.94

Correlation

Comparative studies were done to compare our reagent with ion-selective measurement.

The results from these studies are detailed below.

Correlation coefficient: r=0.9903

Linear regression: y (µmol/l)= 0.999x-4.366

(x= ion selective measurement, y= own reagent).

Specificity

Hemoglobin 77.5 µmol/l (500mg/dl), bilirubin 855 µmol/l (50 mg/dl), lipid 700mg/dl, glucose 55.5 mmol/l (1000mg/dl) and ascorbic acid 1.99 mmol/l (35mg/dl) don't interfere with the assay up to the given levels.

NOTE

In the course of determination please use disposable equipments. Do not use reagents after the expiry date stated on each reagent container label. Do not use products, test solutions and reagents described above for any purpose other than described herein.

For in vitro diagnostic use only.

The following symbols are used on labels

For in vitro diagnostic use

Use by (last day of the month)

Temperature limitation

Batch Code

Code

Bibliography

Schoenfeld, R. G., et al.: Clin. Chem. 10; 533, (1964)